

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT				1. CONTRACT ID CODE <div style="text-align: center;">J</div>		PAGE OF PAGES <div style="text-align: center;">1 2</div>	
2. AMENDMENT/MODIFICATION NO. <div style="text-align: center;">0004</div>		3. EFFECTIVE DATE <div style="text-align: center;">12-Jan-2006</div>		4. REQUISITION/PURCHASE REQ. NO.		5. PROJECT NO. (If applicable)	
6. ISSUED BY NAVAL SURFACE WARFARE CENTER, CARDEROCK CODE 3323, JUNE CATTERTON 9600 MACARTHUR BLVD. WEST BETHESDA MD 20817-5700		CODE <div style="text-align: center;">N00167</div>		7. ADMINISTERED BY (If other than item 6) <div style="text-align: center; font-weight: bold;">See Item 6</div>			
8. NAME AND ADDRESS OF CONTRACTOR (No., Street, County, State and Zip Code)				<input checked="" type="checkbox"/> 9A. AMENDMENT OF SOLICITATION NO. <div style="text-align: center;">N00167-06-Q-5021</div>			
				<input checked="" type="checkbox"/> 9B. DATED (SEE ITEM 11) <div style="text-align: center;">19-Dec-2005</div>			
				10A. MOD. OF CONTRACT/ORDER NO.			
				10B. DATED (SEE ITEM 13)			
CODE				FACILITY CODE			
11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS							
<input checked="" type="checkbox"/> The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of offer <input type="checkbox"/> is extended, <input checked="" type="checkbox"/> is not extended. Offer must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended by one of the following methods: (a) By completing Items 8 and 15, and returning _____ copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or telegram which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by telegram or letter, provided each telegram or letter makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.							
12. ACCOUNTING AND APPROPRIATION DATA (If required)							
13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS. IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.							
A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.							
B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(B).							
C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:							
D. OTHER (Specify type of modification and authority)							
E. IMPORTANT: Contractor <input type="checkbox"/> is not, <input type="checkbox"/> is required to sign this document and return _____ copies to the issuing office.							
14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.) The purpose of this amendment is to provide answers to questions submitted by potential offerors.							
Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.							
15A. NAME AND TITLE OF SIGNER (Type or print)				16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)			
				TEL: _____ EMAIL: _____			
15B. CONTRACTOR/OFFEROR (Signature of person authorized to sign)		15C. DATE SIGNED		16B. UNITED STATES OF AMERICA BY _____ (Signature of Contracting Officer)		16C. DATE SIGNED <div style="text-align: center;">12-Jan-2006</div>	

EXCEPTION TO SF 30
APPROVED BY OIRM 11-24

30-105-04

STANDARD FORM 30 (Rev. 10-83)
Prescribed by GSA
FAR (48 CFR) 53.243

SECTION SF 30 BLOCK 14 CONTINUATION PAGE

SUMMARY OF CHANGES

SECTION SF 1449 - CONTINUATION SHEET

The following have been added by full text:

ADDITIONAL INFORMATION

Relative to the management of dosimeter calibrations:

TAKEN FOR SPECIFICATIONS:

4.1.5.1 Calibration

4.1.5.1.1 Dosimeter shall be initially calibrated by vendor using sources traceable to the National Institute of Standards and Technology (NIST). *[PL1]*

4.1.5.1.2 Dosimeter should deliver the required level of performance / proficiency over its required life expectancy, without requiring recalibration. *[PL2]*

4.1.5.1.3 Calibration should not be affected by removal or replacement of battery / power source. *[PL2]*

4.1.5.2 Periodic Maintenance. Periodic maintenance requirements should be limited to battery replacement and wiping down of exterior surfaces. *[PL2]*

4.1.5.3 Reliability / Life expectancy. Dosimeter should meet all physical, functional, and performance requirements for a period of 10 years minimum, based on a duty cycle of 40,000 hours of normal operation and 10,000 read cycles. *[PL3]*

1. Does the dosimeter require a factory-type calibration to a nationally recognized standard (i.e. NIST) prior to delivery? **Yes, Section 4.1.5.1.1 of the Specification states: "Dosimeter shall be initially calibrated by vendor using sources traceable to the National Institute of Standards and Technology."**

2. Does the Navy have a facility(s) for calibration of the electronic dosimeters? **Section 4.1.5.1.2 of the Specification states: "Dosimeter should deliver the required level of performance/proficiency over its required life expectancy, without requiring recalibration. Also see Section 4.1.5.3."**

3. Does the Navy intend to perform "in-house" calibrations or out-source to a subcontractor/manufacturer? **No, the Navy does not intend to perform "in-house" calibrations or out-source calibrations. See Sections 4.1.5.1.2 and 4.1.5.3.**

4. Does the Navy intend to purchase an automated calibration unit for the electronic dosimeters (Can/should a calibration device be offered)? **No, the Navy does not intend to purchase an automated calibration unit for the electronic dosimeters. See Section 4.1.5.1.2. and 4.1.5.3**